

Eliminate the Risk

# Sani-Cloth<sup>®</sup> Detergent Active



# Sani-Cloth<sup>®</sup> Detergent Active

Activity	Test Laboratory	Standards	Organisms	Clean Cond.	Soiled Cond.	Contact time in mins.
Bactericidal	Lonza AG	EN1040	<i>Pseudomonas aeruginosa</i> , <i>Staphylococcus aureus</i> , <i>Escherichia coli</i> , <i>Enterococcus hirae</i>	X		5
	Dr Brill & Partner GmbH	EN1276	<i>Pseudomonas aeruginosa</i> , <i>Staphylococcus aureus</i> , <i>Escherichia coli</i> , <i>Enterococcus hirae</i>	X	X	5
	The Royal Bolton Hospital	Surface Test	MRSA	X		1
Fungicidal	Dr Brill & Partner GmbH	EN1650	<i>Aspergillus niger</i> , <i>Candida albicans</i>	X	X	15
	Lonza AG	EN1275	<i>Aspergillus niger</i> , <i>Candida albicans</i>	X		15
	Lonza AG	EN13697	<i>Pseudomonas aeruginosa</i> , <i>Staphylococcus aureus</i> , <i>Escherichia coli</i> , <i>Enterococcus hirae</i> , <i>Aspergillus niger</i> , <i>Candida albicans</i>		X	5
Virucidal	MikroLab GmbH	EN14476	<i>Poliovirus Type 1</i> <i>Adenovirus Type 5</i>	X	X	30 60
	Max Von Pettenkofer-Institut/Münich		<i>Hepatitis B</i>	X	X	30
	Microbiological Institute of Milano University		<i>Herpes simplex</i>	X		5
	MikroLab GmbH		<i>Human rotavirus</i>	X		15
	MikroLab GmbH	EN14476 (methodology)	<i>Avian flu</i>	X	X	5 10
	Universita Deigi Studi Di Roma		<i>HIV1</i>	X		30
	Hospital Infection Research Laboratory	EN14476	<i>Poliovirus Type 1</i>	X		30
DGHM Tuberculocidal	Prof. Dr. Med Ralph Schubert, Facharzt für Hygiene und Umweltmedizin		<i>Mycobacterium bovis</i>	X		30



## 225 Count Bucket

Product code : XP00193

Wipe size : 245mmx300mm

Weight : 25g/m<sup>2</sup>

Count : 225 wipes

## Ingredients

Didecylidimethylammonium chloride 0.450 %w/w,

Surfactants 0.325 %w/w, Auxiliary 0.520 %w/w

Excipients 98.705 %w/w

## Alcohol free Detergent/Disinfectant

### Features & Benefits

#### Control

- ✔ Pre-dosed disinfectant wipe providing effective infection control.
- ✔ Mechanical action along with surfactants aids the removal of soil and biofilm which enhances disinfection and provides a log reduction.
- ✔ 14 days residual action safely and effectively eliminates and inhibits growth of micro-organisms (tested organism *Staphylococcus aureus*).

#### Compliance

- ✔ Controlled dosage ensures disinfection for every use.
- ✔ Reduce wastage and risk of errors.
- ✔ Extracted liquid proven effective against organisms tested.

#### Convenience

- ✔ No mixing, no measuring, saves valuable time and storage space.

#### Cost effective

- ✔ Saves valuable time.
- ✔ Competitively priced.

#### Area of use

- ✔ Recommended for use on alcohol sensitive surfaces (see manufacturer's instructions).
- ✔ Used for the disinfection of non-invasive Medical Devices and surfaces.
- ✔ Also used to decontaminate a device prior to sterilisation/high level disinfection.



#### PDI

Aber Park, Aber Road, Flint,

CH6 5EX UK

Tel. +44 1352 736 716

Fax. +44 1352 736 823

www.pdi-europe.co.uk